

Sheet1

TM #1 Electrical Check with Target Attached

	Chassis		60kV Common	
	Ohm Check	Megger Check	Ohm Check	Megger Check
A (Target Oven)	OL	OL	OL	OL
B (Target Oven)				
C (Tube Heater)				
D (Tube Heater)				
E (Gas Lines)				
F (Heat Shield)	OL	931 M $\Omega$	0	0
G (Ext Electrode)	OL	<del>951 M<math>\Omega</math></del> 1.06 M $\Omega$ <del>927 M<math>\Omega</math></del>	OL	OL

	A to G EC		60kv Common to Chassis	
	Ohm Check	Megger Check	Ohm Check	Megger Check
	OL	OL	OL	923 M $\Omega$
	Ohm Check			
	A	B	C	D
A		0	0	0
B			0	0
C				0
D				

UCx #3

November 18, 2011

Megger Checks Done at 500 Volts.