

Vault Target Storage

ISAC Storage Vault Status as of January 5 2012

Tray #	Pail #	Target	Installed (m/d/y)	Reading (mSv/Hr)	Comments	Date Removed	Reading(mSv/Hr)	Comments
1A								
1B								
1C								
2A	87	SiC #25	05/31/11 12/15/11	120mSv/hr 10.7mSv/hr	Not Assayed			Lid has the cable lifting setup
2B	83	SiC #23	10/20/10 12/15/10	143mSv/hr 6.72mSv/hr	Not Assayed			Lid has the cable lifting setup
2C	82	Ta #34	10/20/10 12/14/11	93mSv/hr 4.33mSv/hr	Assayed			Lid has the cable lifting setup
3A								Lid has the cable lifting setup
3B	90	Ta #36	09/29/11	111mSv/hr				Lid has the cable lifting setup
3C	86	UC _x #1	01/14/11 12/12/11	1.7mSv/hr 0.19mSv/hr	Put back in on Jan 5/12 Assayed	01/04/12		Lid has the cable lifting setup
4A	84	SiC #24	12/08/10 11/08/11	27mSv/hr 1.99mSv/hr	Assayed			Lid has the cable lifting setup
4B	76	Nb #4	02/09/10 11/10/11	160mSv/hr 3.89mSv/hr	Assayed			Lid has the cable lifting setup
4C	63	Ta #25	10/02/08 11/10/11	225mSv/hr 4.57mSv/hr	Assayed			Lid has the non cable lifting setup
5A	91	UC _x #2	10/05/11 12/12/11	8.91mSv/hr 1.12mSv/hr	Put back in on Jan 5/12 Assayed	01/04/12		Lid has the cable lifting setup
5B	85	Nb #5	01/11/11	230 mSv/hr				Lid has the cable lifting setup
5C								Lid has the cable lifting setup
<p>All readings are taken at 1m from centre of pail. Sealant (111) is used under lid rim. No contamination found unless other wise noted.</p> <p>All lids have the lever lock ring. This ring must be locked with a thin piece of metal before shipping.</p>								

Vault Target Storage

Tray #	Pail #	Target	Date Installed	Reading (mSv/Hr)	Comments	Date Removed	Reading(mSv/Hr)	Comments
6A	89	Ta#37	14/09/11	840µSv/hr				Lid has the cable
			11/08/11	0.47mSv/hr	Assayed			lifting setup
6B	68	Ta #27	07/14/09	81.0mSv/hr				Lid has the cable
			12/14/11	2.76mSv/hr	Assayed			lifting setup
6C	88	Ta#35	07/25/11	411mSv/hr				Lid has the cable
								lifting setup
7A	75	Ta#31	06/10/10	116mSv/hr				Lid has the cable
			12/15/11	4.17mSv/hr	Assayed			lifting setup
7B	92	TaC#1	11/16/11	239mSv/hr				Lid has the cable
								lifting setup
7C	81	ZrC#5	08/25/10	292mSv/hr				Lid has the cable
			12/15/11	13.7mSv/hr	Not Assayed			lifting setup
8A	64	Ta #26	02/06/09	131mSv/hr				Lid has the non cable
			12/14/11	3.37mSv/hr	Assayed			lifting setup
8B	78	Ta#32	07/16/10	124mSv/hr				Lid has the cable
								lifting setup
8C	79	SiC#22	07/16/10	70.1mSv/hr				Lid has the cable
			12/14/11	1.93mSv/hr	Assayed			lifting setup
Tray #	Description of Item	Date Installed	Reading (mSv/Hr)	Comments	Date Removed	Reading(mSv/Hr)	Comments	
9	Used source tray	01/15/09	5.6 to 16.4					
	from TM #2		mSv/hr thru					
			360°					
10	Used source tray	05/12/11	1.66 to 3.50	The source tray was				
	from TM #1		mSv/hr thru	taken apart for inspection				
			360°					
All readings are taken at 1m from centre of pail. Sealant (111) is used under lid rim. No contamination found unless other wise noted.								
All lids have the lever lock ring. This ring must be locked with a thin piece of metal before shipping.								

Vault Target Storage

Vault Target Storage