

2013-08-15

TM4 ITE LEAK CHECK

Leak Detector

Base leak rate: 1.1×10^{-7} atm cc/s

Base pressure: 5.9×10^{-2} torr

pressurize helium water lines: 70 psi helium

HS: 3min No response.

Coil 2: 1:46:30 started, 1:49:44 finished

Coil 1: got response immediately to
leak rate maximum scale
pressure up to -1 torr range
all High Vacuum gauges got response.

Tube Heater (+)

Tube Heater (-)