

Target Module with Target Check		
Document-75583	Release No. 1	Release Date.: 2013/12/11

## 4.2 TM3/4SIS with Target and No External Connections

Shunts installed and tightened E-F-G-H-T-J-RFQ(H)	
FEBIAD coil shunts installed and tightened P-Q-R-U	

	Ohm Meter Check				
	Target (ABCD)	Einzel Lens (Y)	60kV Com (T)	Chassis	EE (I)
Magnet Coil (U, Q)					
Anode (H)					
60kV Common (T)	OL	OL			
	Infinite	Infinite			
Chassis	OL	OL	OL		
	Infinite	Infinite	Infinite		
EE (I)	OL	OL	OL	OL	
	Infinite	Infinite	Infinite	Infinite	
Einzel Lens (Y)	OL				
	Infinite				

	Megger Check (all @ 500V unless otherwise noted)				
	Target (ABCD)	Einzel Lens (Y)	60kV Com (T)	Chassis	EE (I)
Anode (H)					
60kV Common (T)					
Chassis	2.09M	OL			
	Infinite @ 1000V	Infinite @ 1000V			
EE (I)	208.5M	OL	203.6M		
	Infinite @ 1000V	Infinite @ 1000V	Infinite		
Einzel Lens (Y)	OL				
	Infinite				

Conductance Check [mΩ]						
A - B	A - C	A - D	B - C	B - D	C - D	
8.070m	7.290	5.609	8.028	6.363	5.516	
4 - 5.7	5 - 6.8	4 - 5.3	5 - 6.8	4 - 5.3	4 - 5.5	

Backg. leak rate:		HS leak rate:	
[atm-cc/s]	< 5E-9	[atm-cc/s]	< Backg. leak rate

Target: UC #10	Date: Sept 23 Sep 2014
Location: SHC CS SHCpost	Signature: T LAVE