

## 4.2 TM3/4 SIS with Target and No External Connections

Shunts installed and tightened E-F-G-H-T-J-RFQ(H)	✓
FEBIAD coil shunts installed and tightened P-Q-R-U	✓

Ohm Meter Check					
	Target (ABCD)	Einzel Lens (Y)	60kV Com (T)	Chassis	EE (I)
Magnet Coil (U, Q)					
Anode (H)					
60kV Common (T)	OL Infinite	OL Infinite			
Chassis	OL Infinite	OL Infinite			
EE (I)	OL Infinite	OL Infinite	OL Infinite	OL Infinite	
Einzel Lens (Y)	OL Infinite				

Megger Check (all @ 500V unless otherwise noted)					
	Target (ABCD)	Einzel Lens (Y)	60kV Com (T)	Chassis	EE (I)
Anode (H)					
60kV Common (T)					
Chassis	168.3M Infinite @ 1000V	1.0G Infinite @ 1000V	167M Infinite		
EE (I)	707K Infinite	1.0G Infinite	56K Infinite		
Einzel Lens (Y)	∞ Infinite				

Conductance Check [mΩ]						
A - B	A - C	A - D	B - C	B - D	C - D	
4.741	5.581	4.419	5.217	4.042	4.842	
4 - 5.7	5 - 6.8	4 - 5.3	5 - 6.8	4 - 5.3	4 - 5.5	

Backg. leak rate:	HS leak rate:
[atm-cc/s]	[atm-cc/s]
< 5E-9	< Backg. leak rate

Target: <u>Ta #45</u>	Date: <u>June 2015</u>
Location: SHC CS <u>SHCpost</u>	Signature: <u>T CAVE</u> <u>D WANG</u>