

Vault Target Storage

ISAC Storage Vault Status as of November 26 2015

Tray #	Pail #	Target	Initial	Second	Third	Status	Comments	Removed (m/d/y)
			Reading (mSv/Hr)	Reading (mSv/Hr)	Reading (mSv/Hr)			
<b>1A</b>	128	UC <sub>x</sub> #13	21.5mSv/hr					
Date(m/d/y)			09/24/15					
<b>1B</b>								
Date(m/d/y)								
<b>1C</b>	125	UC <sub>x</sub> #12	25.1mSv/hr	5.12mSv/hr		Assayed		
Date(m/d/y)			07/02/15	08/04/15				
<b>2A</b>	RH B/L	2A3 Window	8mSv/hr	0.81mSv/hr	0.494mSv/hr	Assayed	Unmodified lid, no sealant	
Date(m/d/y)			04/30/14	05/14/14	08/05/15			
<b>2B</b>	122	ThO#1	6.20mSv/hr	1.44mSv/hr		Assayed		
Date(m/d/y)			03/13/15	08/05/15				
<b>2C</b>	117	SiC #30	211mSv/hr	34.1mSv/hr	6.83mSv/hr	Measured		
Date(m/d/y)			08/06/14	09/15/14	08/05/15			
<b>3A</b>	96	SiC #27	178mSv/hr	3.66mSv/hr	2.70mSv/hr	Assayed		
Date(m/d/y)			07/06/12	05/15/14	09/15/14			
<b>3B</b>	129	ZrC #7	157mSv/hr				Pail contains copper water line TM#4	
Date(m/d/y)			10/23/15					
<b>3C</b>	127	SiC #31	489mSv/hr					
Date(m/d/y)			07/30/15					
<b>4A</b>	121	Ta #44	276mSv/hr	18.0mSv/hr		Measured		
Date(m/d/y)			11/21/14	08/06/15				
<b>4B</b>	124	Ta #45	218mSv/hr	42.9mSv/hr		Measured	Contamination may be present on the clasp.	
Date(m/d/y)			06/04/15	08/06/15				
<b>4C</b>	106	Ta #39	95.3mSv/hr	6.06mSv/hr	4.57mSv/hr	Assayed	Measured 08/06/15 2.93mSv/hr	
Date(m/d/y)			07/02/13	05/21/14	09/12/14			
<b>5A</b>	107	Ta #40	224mSv/hr	15.6mSv/hr	4.93mSv/hr	Measured		
Date(m/d/y)			07/29/13	05/16/14	08/05/15			
<b>5B</b>	115	TiC #4	407mSv/hr	9.12mSv/hr		Measured		
Date(m/d/y)			06/27/14	08/05/15				
<b>5C</b>	120	ZrC #6	112mSv/hr	7.38mSv/hr		Measured		
Date(m/d/y)			10/21/14	08/05/15				

Initial readings are taken at 1m from centre of pail. Assayed readings are taken at 1m from side of pail. Sealant (111) is used under lid rim.

All lids have the lever lock ring and cable lifting setup, ring is locked with a thin piece of metal before shipping. All pails are checked for contamination

Vault Target Storage

Tray #	Pail #	Target	Initial	Second	Third	Status	Comments	Removed (m/d/y)
			Reading (mSv/Hr)	Reading (mSv/Hr)	Reading (mSv/Hr)			
<b>6A</b>								
Date(m/d/y)								
<b>6B</b>	126	Ta #46	265 mSv/hr				Can contains various target bits from	
Date(m/d/y)			11/25/15				the HC clean up.	
<b>6C</b>	88	Ta #35	411mSv/hr	8.26mSv/hr	4.69mSv/hr	Assayed	Measured 09/12/14 4.70mSv/hr	
Date(m/d/y)			07/25/11	06/20/13	05/21/14		Measured 08/06/15 3.5mSv/hr	
<b>7A</b>	112	Ta #41	536mSv/hr	37.6mSv/hr	6.59mSv/hr	Measured	Pail is unnumbered	
Date(m/d/y)			12/09/13	05/16/14	08/05/15			
<b>7B</b>	114	Ta #42	277mSv/hr	5.17mSv/hr		Measured	Metal grabby tab is on lid	
Date(m/d/y)			05/30/14	08/05/15				
<b>7C</b>	118	Ta #43	150mSv/hr	5.48mSv/hr		Measured		
Date(m/d/y)			09/05/14	08/05/15				
<b>8A</b>								
Date(m/d/y)								
<b>8B</b>								
Date(m/d/y)								
<b>8C</b>								
Date(m/d/y)								
Tray #	Description of Item	Initial	Second	Third	Status	Comments	Removed (m/d/y)	
		Reading (mSv/Hr)	Reading (mSv/Hr)	Reading (mSv/Hr)				
9	Used source tray from TM #2	5.6-16.4mSv/hr though 360°	2.07-4.39mSv/hr though 360°		Measured			
Date(m/d/y)		01/15/09	11/02/12					
10	Used source tray from TM #1	1.66-3.50mSv/hr though 360°	1.83-3.23 mSv/hr though 360°		Measured	The source tray was taken apart for inspection		
Date(m/d/y)		05/12/11	11/02/12					
Initial readings are taken at 1m from centre of pail. Assayed readings are taken at 1m from side of pail. Sealant (111) is used under lid rim.								
All lids have the lever lock ring and cable lifting setup, ring is locked with a thin piece of metal before shipping. All pails are checked for contamination								

## Vault Target Storage

## Vault Target Storage