

~~MSP - Temp A - OK~~
~~Had - Temp A - OK~~

MSP - Heat Shield - Gas line - are tied together
 P-Q - tied together
 R+U - tied together
 IN containment Box
 Target Module Electrical Check with no target attached for Modules 2/3/4

	A (Target)	B (Target)	C (Target)	D (Target)	Q (Coil)	R (Coil)	T(60kV Common)	X (55 Pin Con)	I (EE)	Y (Einzel Lens)	Chassis
MSP											
A (Target)	OK										
B (Target)	OK	OK									
C (Target)	OK	OK	OK								
D (Target)	OK	OK	OK	GL							
P-Q (Coil)	OK	OK	OK	OK	OK						
U-R (Coil)	OK	OK	OK	OK	OK						
T(60kV Common)	OK	OK	OK	OK	OK	OK					
X(55 Pin con)	OK	OK	OK	OK	OK	OK	GL				
I (EE)	OK	OK	OK	OK	OK	OK	OK	OK			
Y (Einzel Lens)	OK	OK	OK	OK	OK	OK	OK	OK	OK		
Chassis	OK	OK	OK	OK	OK	OK	OK	OK	OK	OK	

Meggering check all circuits checked at 500V. All values should be infinite. Any measured values are in ohms.

Module #	2	3	4	Date:	April 12 2017
Location:	SHC	CS	Other:	Signatures:	D. Wang T. CAVE