

Vault Target Storage

ISAC Storage Vault Status as of October 23 2017

Tray #	Pail #	Target	Initial	Second	Third	Status	Comments	Removed (m/d/y)
			Reading (mSv/Hr)	Reading (mSv/Hr)	Reading (mSv/Hr)			
1A	159	UC#21	27.4mSv/hr					
Date(m/d/y)			10/23/17					
1B	150	UC#20	30.1mSv/hr					
Date(m/d/y)			06/29/17					
1C	145	UC#19	22.8mSv/hr					
Date(m/d/y)			05/18/17					
2A								
Date(m/d/y)								
2B	141	SiC#33	295mSv/hr	25.1mSv/hr		Measured		
Date(m/d/y)			08/25/16	03/20/17				
2C	117	SiC #30	211mSv/hr	34.1mSv/hr	6.83mSv/hr	Assayed	4.70mSv/hr June 30 th 2016	
Date(m/d/y)			08/06/14	09/15/14	08/05/15		2.83mSv/hr 03/20/17	
3A	132	SiC #32	79.7mSv/hr	24.5mSv/hr	8.21mSv/hr	Measured		
Date(m/d/y)			03/22/16	06/30/16	03/20/17			
3B	129	ZrC #7	157mSv/hr	15.2mSv/hr	3.85mSv/hr	Assayed	Pail contains copper water line TM#4	
Date(m/d/y)			10/23/15	06/30/16	03/20/17			
3C	127	SiC #31	489mSv/hr	17.5mSv/hr	7.90mSv/hr	Measured		
Date(m/d/y)			07/30/15	06/30/16	03/20/17			
4A	121	Ta #44	276mSv/hr	18.0mSv/hr	8.44mSv/hr	Assayed	4.37mSv/hr 03/20/17	
Date(m/d/y)			11/21/14	08/06/15	06/30/16			
4B	124	Ta #45	218mSv/hr	42.9mSv/hr	13.1mSv/hr	Assayed	Contamination may be present on the Clasp. 5.72mSv/hr 03/20/17	
Date(m/d/y)			06/04/15	08/06/15	06/30/16			
4C	137	Ta #47	114mSv/hr	42.5mSv/hr	5.92mSv/hr	Measured		
Date(m/d/y)			06/15/16	06/30/16	03/20/17			
5A	151	SiC #36	190mSv/hr					
Date(m/d/y)			08/03/17					
5B	115	TiC #4	407mSv/hr	9.12mSv/hr	5.12mSv/hr	Assayed	3.18mSv/hr 03/20/17	
Date(m/d/y)			06/27/14	08/05/15	06/30/16			
5C	164	Ta #53	244mSv/hr					
Date(m/d/y)			09/28/17					
Initial readings are taken at 1m from centre of pail. Assayed readings are taken at 1m from side of pail. Sealant (111) is used under lid rim.								
All lids have the lever lock ring and cable lifting setup, ring is locked with a thin piece of metal before shipping. All pails are checked for contamination								

Vault Target Storage

Tray #	Pail #	Target	Initial	Second	Third	Status	Comments	Removed (m/d/y)
			Reading (mSv/Hr)	Reading (mSv/Hr)	Reading (mSv/Hr)			
6A	146	Ta #49	134 mSv/hr	19.3mSv/hr		Measured	Lid should be checked for contamination	
Date(m/d/y)			10/20/16	03/20/17				
6B	126	Ta #46	265 mSv/hr	22mSv/hr	7.31mSv/hr	Measured	Can contains various target bits from the HC clean up.	
Date(m/d/y)			11/25/15	06/30/16	03/20/17			
6C	139	Ta #48	201 mSv/hr	17.2mSv/hr		Measured		
Date(m/d/y)			07/11/16	03/20/17				
7A	112	Ta #41	536mSv/hr	37.6mSv/hr	6.59mSv/hr	Assayed	Pail is unnumbered. 5.31mSv/hr June 30 th 2016	
Date(m/d/y)			12/09/13	05/16/14	08/05/15		3.68mSv/hr 03/20/17	
7B	143	SiC #34	125mSv/hr	8.17mSv/hr		Measured		
Date(m/d/y)			11/03/16	03/20/17				
7C	144	SiC #35	228mSv/hr	18.50Sv/hr		Measured		
Date(m/d/y)			11/24/16	03/20/17				
8A	152	Ta #52	38.1mSv/hr					
Date(m/d/y)			08/30/17					
8B	149	Ta #51	144mSv/hr					
Date(m/d/y)			06/12/17					
8C	148	Ta #50	38.2mSv/hr	16.3mSv/hr		Measured		
Date(m/d/y)			01/26/17	03/20/17				
Tray #	Description of Item	Initial	Second	Third	Status	Comments	Removed (m/d/y)	
9	Used source tray from TM #2	5.6-16.4mSv/hr though 360°	2.07-4.39mSv/hr though 360°	0.50-1.18mSv/hr though 360°	Measured			
Date(m/d/y)		01/15/09	11/02/12	10/12/17				
10	Used source tray from TM #1	1.66-3.50mSv/hr though 360°	1.83-3.23 mSv/hr though 360°	0.52-1.15mSv/hr though 360°	Measured	The source tray was taken apart for inspection		
Date(m/d/y)		05/12/11	11/02/12	10/12/17				
Initial readings are taken at 1m from centre of pail. Assayed readings are taken at 1m from side of pail. Sealant (111) is used under lid rim.								
All lids have the lever lock ring and cable lifting setup, ring is locked with a thin piece of metal before shipping. All pails are checked for contamination								

Vault Target Storage

Vault Target Storage