

Loc	Target	Pail	Assay 1 [mSv/h]	A1 date	Assay 2 [mSv/h]	A2 date	Assay 3 [mSv/h]	A3 date	Assay 4 [mSv/h]	A4 date	Assay 5 [mSv/h]	A5 date	Assayed	Comment
1A	UCR38	274	33.3	9/15/2022	4.8	12/16/2022							0	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope. contains one proton beam window without copper cooling lines in addition to the target
1B	UCR40	278	15.6	1/11/2023									0	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
1C	UCR41	231	14.0	1/30/2023									0	Pail is likely contaminated, Contains two proton beam window without copper cooling lines in addition to the target This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
2A	UCR36	265	19.6	1/12/2022	10.7	1/28/2022							Measured	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
2B	SICR42	271	9.1	7/6/2022									0	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
2C	CR3	270	37.9	11/10/2021	10.8	1/28/2022							Measured	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
3A	TICR5	279	40.0	6/21/2023									0	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
3B	TAR63	266	410.0	20/01/2022									0	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
3C	SICR45	280	138.0	7/27/2023									0	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
4A	SICR44	275	42.6	10/13/2022									0	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
4B	UCR39	276	14.4	11/3/2022									0	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
4C	CR4	277	164.0	11/24/2022									0	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
5A	SICR43	272	47.8	28/07/2022									0	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
5B	SICR41	260	33.1	13/10/2021	7.8	1/30/2022	1.5	12/16/2022					Measured	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
5C	TAR64	273	86.6	8/18/2022									0	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
6A	TICR6	232	151.0	8/23/2023									0	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
6B	UCR42	282	34.9	9/1/2023									0	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
6C	NBR9	246	149.0	1/21/2021	9.1	1/31/2022	2.0	12/16/2022					Measured	This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
7A	TAR62	256	58.0	9/17/2021	16.9	1/27/2022	3.6	2022-12-16					Measured	2022-12-16 moved from 7A to 7B in vault. AN Pail placed back in 7A - TC Disc in bottom of pail was not installed pail will need repacking. This pail is likely contaminated and also contains 2A2M19. This pail is of previous revision, not certified as lifting device. Prior to shipping, a 3" loop must be crimped at the end of the wire rope.
7B	NBR11	283	65.7	10/6/2023									0	0
8C														
8A														
8B														
8C														
9	TM4 optics tray 2018, TM2 source tray 2019		1.9	6/14/2018	14.0	3/1/2019	7.1	11/2/2020						In Source Tray Box. The 1.9 mSv/hr value is just for the optics tray. All other values are for both Optics and Source tray.
10	TM4 Source tray		22.1	6/10/2019										7.5 to 22.1 mSv/hr on 360 rotation.

Need to change pail to new design